

# GIBSON LABORATORIES, LLC

1040 MANCHESTER STREET  
LEXINGTON, KY 40508

Telephone (859) 254 - 9500

Fax (859) 253 - 1476

## QUALITY CONTROL CERTIFICATE

### I. PRODUCT INFORMATION

**72106 FF INOCULATING FLUID**

**LOT NUMBER: 1195061**

**EXP DATE: 07/12/2012**

Representative samples have been tested by the Quality Control Laboratory.

### II. BIOLOGICAL PERFORMANCE

Procedures: Standardized aliquots of the listed ATCC cultures were inoculated onto selected samples for performance testing. Incubation times range from 24-48 hours at 35-37C for bacteria, and 3-5 days at 25 - 30C for fungi.

Samples were examined for physical characteristics listed below.

Organisms	ATCC	Growth/Reaction
<i>Nonapplicable</i>		

Note: Quality Control procedures involving this product meets applicable CLSI (M22 ) Standards

### III. PHYSICAL CHARACTERISTICS

**APPEARANCE:** Clear Liquid

**GEL STRENGTH:** N/A

**BIOBURDEN:** Inspected after 3 days at 35C: Negative for Bacterial Contamination  
Inspected after 7 days at 25C: Negative for Fungal Contamination

**FINAL pH:** N/A

*Jerry Stephenson*

Quality Control