

GIBSON LABORATORIES, LLC

1040 MANCHESTER STREET
LEXINGTON, KY 40508

Telephone (859) 254 - 9500
Fax (859) 253 - 1476

QUALITY CONTROL CERTIFICATE

I. PRODUCT INFORMATION

72101 GN/GP INOCULATING FLUID
LOT NUMBER: 1123025
EXP DATE: 05/01/2012

Representative samples have been tested by the Quality Control Laboratory.

II. BIOLOGICAL PERFORMANCE

Procedures: Standardized aliquots of the listed ATCC cultures were inoculated onto selected samples for performance testing. Incubation times range from 24-48 hours at 35-37C for bacteria, and 3-5 days at 25 - 30C for fungi. Samples were examined for physical characteristics listed below.

| Organisms | ATCC | Growth/Reaction |
|-----------------------|------|-----------------|
| <i>Non-applicable</i> | | |

Note: Quality Control procedures involving this product meets applicable CLSI (M22) Standards

III. PHYSICAL CHARACTERISTICS

APPEARANCE: Clear liquid
GEL STRENGTH: N/A
BIOBURDEN: Inspected after 3 days at 35C: Negative for Bacterial Contamination
Inspected after 7 days at 25C: Negative for Fungal Contamination
FINAL pH: N/A

Jerry Stephenson

Quality Control