

# GIBSON LABORATORIES, LLC

1040 MANCHESTER STREET  
LEXINGTON, KY 40508

Telephone (859) 254 - 9500  
Fax (859) 253 - 1476

## QUALITY CONTROL CERTIFICATE

### I. PRODUCT INFORMATION

**72101 GN/GP INOCULATING FLUID**  
**LOT NUMBER: 1125039**  
**EXP DATE: 05/03/2012**

Representative samples have been tested by the Quality Control Laboratory.

### II. BIOLOGICAL PERFORMANCE

Procedures: Standardized aliquots of the listed ATCC cultures were inoculated onto selected samples for performance testing. Incubation times range from 24-48 hours at 35-37C for bacteria, and 3-5 days at 25 - 30C for fungi. Samples were examined for physical characteristics listed below.

Organisms	ATCC	Growth/Reaction
<i>Non-applicable</i>		

Note: Quality Control procedures involving this product meets applicable CLSI (M22 ) Standards

### III. PHYSICAL CHARACTERISTICS

**APPEARANCE:** Clear liquid  
**GEL STRENGTH:** N/A  
**BIOBURDEN:** Inspected after 3 days at 35C: Negative for Bacterial Contamination  
Inspected after 7 days at 25C: Negative for Fungal Contamination  
**FINAL pH:** N/A

*Jerry Stephenson*  
Quality Control