

GIBSON LABORATORIES, LLC

1040 MANCHESTER STREET
LEXINGTON, KY 40508

Telephone (859) 254 - 9500

Fax (859) 253 - 1476

QUALITY CONTROL CERTIFICATE**I. PRODUCT INFORMATION**

72101 GN/GP INOCULATING FLUID

LOT NUMBER: 1227003

EXP DATE: 08/13/2012

Representative samples have been tested by the Quality Control Laboratory.

II. BIOLOGICAL PERFORMANCE

Procedures: Standardized aliquots of the listed ATCC cultures were inoculated onto selected samples for performance testing. Incubation times range from 24-48 hours at 35-37C for bacteria, and 3-5 days at 25 - 30C for fungi. Samples were examined for physical characteristics listed below.

Organisms	ATCC	Growth/Reaction
<i>Non-applicable</i>		

Note: Quality Control procedures involving this product meets applicable CLSI (M22) Standards

III. PHYSICAL CHARACTERISTICS

APPEARANCE: Clear liquid

GEL STRENGTH: N/A

BIOBURDEN: Inspected after 3 days at 35C: Negative for Bacterial Contamination
Inspected after 7 days at 25C: Negative for Fungal Contamination

FINAL pH: N/A

Jerry Stephenson
Quality Control