



21124 Cabot Blvd.  
 Hayward, California 94545  
 510-785-2564 Telephone  
 800-284-4949 Orders  
 510-782-4639 Fax  
 www.biolog.com

**CERTIFICATE OF PERFORMANCE**

Product No.: 80102  
 Name: RAINBOW® AGAR 0157  
 Lot No.: 912091  
 Expiration Date: 12/12

**QUALITY PROFILE**

Test 1. Efficiency of plating compared to TSA+B medium

Rainbow O157 comparable at:  
 10 ul of approx. 10<sup>3</sup> cells/ml Passes test

Test 2. Colony coloration of 16 bacterial strains

|                             |                |             |
|-----------------------------|----------------|-------------|
| Species tested              |                |             |
| E. coli                     | K. pneumoniae  |             |
| E. coli O157 (2 strains)    | E. amnigenus   |             |
| E. coli O111:H8 (2 strains) | S. sonnei      |             |
| E. coli O26:H11 (2 strains) | P. aeruginosa  |             |
| E. coli O48:H21             | E. faecalis    |             |
| H. alvei (2 strains)        | S. epidermidis | Passes test |

Test 3. Verification of Indole Spot Test

|                 |          |             |
|-----------------|----------|-------------|
| Species tested  |          |             |
| E. coli         | positive |             |
| E. coli O157:H7 | positive |             |
| H. alvei A      | negative |             |
| P. aeruginosa   | negative | Passes test |

Samples of this lot have met performance standards established by Biolog, Inc.

Technician: \_\_\_\_\_ Completion date: \_\_\_\_\_

**INTENDED USE**

**NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.**



31 December 2009

APPROVAL: QUALITY ASSURANCE  
 Barbara B. So

DATE