



21124 Cabot Blvd.
 Hayward, California 94545
 510-785-2564 Telephone
 800-284-4949 Orders
 510-782-4639 Fax
 www.biolog.com

CERTIFICATE OF PERFORMANCE

Product No.: 71007
 Name: BUA Agar W/5% SHEEP BLOOD
 Lot No.: PBUS11206
 Expiration Date: 06 JUN 2012

QUALITY PROFILE

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>	<u>Identification</u>
Bifidobacterium breve	ATCC 15700	Good Growth	Identifies
Clostridium sordellii	ATCC 9714	Good Growth	Identifies
Lactobacillus casei	ATCC 393	Good Growth	Identifies
Peptostreptococcus micros	ATCC 33270	Good Growth	Identifies

Final pH: 7.2 ± 0.2 @ 25° C.
 Unit: Plate
 Volume: 23 ml

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.



30 December 2011

APPROVAL: QUALITY ASSURANCE
 Barbara B. So

DATE