

GIBSON LABORATORIES, LLC

1040 MANCHESTER STREET
LEXINGTON, KY 40508

Telephone (859) 254 - 9500

Fax (859) 253 - 1476

QUALITY CONTROL CERTIFICATE**I. PRODUCT INFORMATION**

71106 2% MALT EXTRACT
LOT NUMBER: 1224092
EXP DATE: 11/04/11

Representative samples have been tested by the Quality Control Laboratory.

II. BIOLOGICAL PERFORMANCE

Procedures: Standardized aliquots of the listed ATCC cultures were inoculated onto selected samples for performance testing. Incubation times range from 24-48 hours at 35-37C for bacteria, and 3-5 days at 25 - 30C for fungi. Samples were examined for physical characteristics listed below.

Organisms	ATCC	Growth/Reaction
<i>Candida albicans</i>	10231	Growth within 7 days
<i>Trichophyton mentagrophytes</i>	9533	Growth within 7 days

Note: Quality Control procedures involving this product meets applicable CLSI (M22) Standards.

III. PHYSICAL CHARACTERISTICS

APPEARANCE: Light tan - clear
GEL STRENGTH: Firm
BIOBURDEN: Inspected after 3 days at 35C: Negative for Bacterial Contamination
FINAL pH: 5.5+/- 0.2

Jerry Stephenson
Quality Control