

# BiOLOG

21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

## CERTIFICATE OF PERFORMANCE

Product No.: 72232  
Name: PM IFY-10 Base (1.2x)  
Lot No.: 1108231  
Expiration Date: 23 AUG 2012

---

### QUALITY PROFILE

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

NO GROWTH PROMOTION REQUIRED

---

Final pH:	6.0 ± 0.2 @ 25° C.
Color:	Colorless, slightly turbid liquid
Unit:	bottle
Volume:	125 ml

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35°C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

---

### INTENDED USE

FOR RESEARCH USE ONLY.

---



26 September 2011

APPROVAL: QUALITY ASSURANCE  
Barbara B. So

DATE

CA72232  
May 2008