



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF ANALYSIS

Product No.: 3415  
Name: YT Turbidity Standard  
Lot No.: Y122718  
Expiration Date: 27 SEP 2020

---

### QUALITY PROFILE

---

#### Results

Standard check:  
47.0 ± 0.5 % transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

---

### STORAGE RECOMMENDATIONS

Store in the dark at 20 - 26 °C.

---

### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

---

  
APPROVAL: QUALITY ASSURANCE

04 Jan 2019  
DATE