



21124 Cabot Blvd.
Hayward, California 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF ANALYSIS

Product No.: 3415
Name: YT Turbidity Standard
Lot No.: Y062619
Expiration Date: 26 MAR 2021

QUALITY PROFILE

	Results
Standard check: 47.0 ± 0.5 % transmittance	Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

STORAGE RECOMMENDATIONS

Store in the dark at 20° - 26°C.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.


APPROVAL: QUALITY ASSURANCE

06/27/2019
DATE