



CERTIFICATE OF ANALYSIS

Product Description: Biolog IF-M2 (Product# 72302)

Catalog Number: R-071-C

Lot Number: 71017-1

Date of Manufacture: 17-Oct-2017

Expiration Date: 17-Oct-2020

Storage Temperature: 2 - 8°C

Storage Instructions: Protect from direct light

For Kit Products the individual component products and their respective Lot numbers are to be listed below:

Component Product Code:	Component Lot Number:
N/A	N/A
N/A	N/A
N/A	N/A
N/A	N/A

CAUTION: Not for diagnostic use. This product is for Research Use Only.

This lot conforms to specifications established by Millipore (UK) LTD for this product.

Note: Where N/A has been inserted in the table below no testing has been performed, as per the pertinent Product Specification, on this property / characteristic

Test or Property / Characteristic	Specification	Result
Appearance	Clear with a Faint Yellow Tint	Clear with a Faint Yellow Tint
Quality Control Acceptance Date	N/A	01-Nov-2017
Sterility Testing	PASS or N/A	PASS
pH	7.0 – 7.3	7.0
Osmolality	266 – 269	268
ES Qualification	N/A	N/A
Plating Efficiency	Cells detached < 10 minutes and replaced with > 85% efficiency	PASS
Freeze Thaw	N/A	N/A
Endo Toxin	N/A	N/A
Mycoplasma	N/A	N/A
3 day Embryonic Cell Development Test *	N/A	N/A

* Test Results as per complete media product

Martin Ross

Quality Assurance Signature

20171102

Date

Millipore (UK) Ltd
2-4 Fleming Road
Kirkton Campus
Livingston
West Lothian-EH547BN
UK

www.millipore.com