



## Certificate of Analysis

Catalog No.: 72231  
Product Name: PM IFY-0 Base (1.2X)  
Lot No.: 1908071  
Expiration Date: 07 AUG 2020

---

### Quality Profile

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
NO GROWTH PROMOTION REQUIRED		

---

Final pH: 6.0 ± 0.2 at 25 °C.  
Color: Colorless, slightly turbid liquid  
Unit: Bottle  
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

---

### Intended Use

---

For research use only.

  
Quality Assurance Approval

08/28/2019  
Date



21124 Cabot Blvd  
Hayward, CA 94545 USA

Tel: +510 785 2564  
Fax: +510 782 4639

[www.biolog.com](http://www.biolog.com)

CA72231  
Rev B  
Effective Date: 05/23/2019