



21124 Cabot Blvd.  
Hayward, CA 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF PERFORMANCE

**Catalog No.:** 80102  
**Product Name:** Rainbow® Agar O157 Dehydrated Media  
**Lot No.:** 910151  
**Expiration Date:** 30 NOV 2022

#### QUALITY PROFILE

**Test 1:** Efficiency of plating compared to TSA+B medium.

**Result:** Pass

**Test 2:** Colony coloration of 16 bacterial strains.

Quality Control Organism	Source	Expected Color	Result
<i>Escherichia coli</i>	ATCC 11775	Red/Magenta	Pass
<i>Escherichia coli</i> O157:H7	USDA/MD O157/H7	Charcoal/Black	Pass
<i>Escherichia coli</i> O157:H7 GLUC+	CDC 5101	Dark Blue	Pass
<i>Escherichia coli</i> O111:H8	GEULPH 517	Blue/Purple	Pass
<i>Escherichia coli</i> O111:H8	GEULPH 541	Blue/Purple	Pass
<i>Escherichia coli</i> O26:H11	GEULPH 561	Blue/Red	Pass
<i>Escherichia coli</i> O26:H11	GEULPH 573	Dark Blue	Pass
<i>Escherichia coli</i> O48:H21	PHL-BERKELEY 94A 9057	Dark Blue	Pass
<i>Hafnia alvei</i>	CCUG 429	Blue/Charcoal	Pass
<i>Hafnia alvei</i>	WAUTERS-UCDL 2	Colorless	Pass
<i>Klebsiella pneumoniae</i> ss <i>pneumoniae</i>	WAUTERS-UCDL 101	Light Blue/Orange	Pass
<i>Enterobacter amnigenus</i>	ATCC 33072	Colorless	Pass
<i>Shigella sonnei</i>	MUTTERS-HIM 619-4	Blue Center	Pass
<i>Pseudomonas aeruginosa</i>	ATCC 10145	Light Yellow/Green	Pass
<i>Enterococcus faecalis</i>	ATCC 19433	Colorless	Pass
<i>Staphylococcus epidermidis</i>	VICTORIA HOSP 1957	Colorless	Pass

**Test 3:** Verification of Indole Spot Test.

**Result:** Pass

Samples of this lot have performance standards established by Biolog, Inc.

**Technician:** David Brown **Completion Date:** 11/08/2019

#### INTENDED USE

Not for human *in vitro* diagnostic use.

Asipath  
**Quality Assurance Approval**

11/08/2019  
**Date**