



21124 Cabot Blvd.
Hayward, California 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF PERFORMANCE

Product No.: 72401
Name: Inoculating Fluid -A (IF-A)
Lot No.: 20FEB061
Expiration Date: 06 FEB 2021

QUALITY PROFILE

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
NO GROWTH PROMOTION REQUIRED		

Final pH: 7.1 ± 0.2 at 25 °C.
Color: Light Green
Unit: tube
Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.


APPROVAL: QUALITY ASSURANCE

10 Mar 2020
DATE