



## Certificate of Analysis

Catalog No.: 1005  
Product Name: YT MicroPlate™  
Lot No.: 3502231  
Expiration Date: 23 NOV 2021

---

### Quality Profile

---

#### Results

Substrate position check: Passes test  
Substrate response check: Passes test  
ID performance check:  
Contamination check  
(No growth at 96 hours): Passes test

---


#### Storage Recommendations:

Store between 2 °C and 8 °C

---

#### Intended Use

Not for Human *in vitro* diagnostic use

  
Quality Assurance

22 Mar 2021  
Date

CA1005 Rev C  
Effective Date: 12/20/2019



21124 Cabot Blvd  
Hayward, CA 94545 USA

Tel: +510 785 2564  
Fax: +510 782 4639

www.biolog.com