



21124 Cabot Blvd.
Hayward, California 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF ANALYSIS

Product No.: 3427
Name: AN Turbidity Standard
Lot No.: A033020
Expiration Date: 30 DEC 2021

QUALITY PROFILE

Results

Standard check:
65.0 ± 0.5 % transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

STORAGE RECOMMENDATIONS

Store in the dark at 20° - 26°C.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.


APPROVAL: QUALITY ASSURANCE

08 Apr 2020
DATE