



21124 Cabot Blvd.  
 Hayward, California 94545  
 510-785-2564 Telephone  
 800-284-4949 Orders  
 510-782-4639 Fax  
 www.biolog.com

**CERTIFICATE OF PERFORMANCE**

Product No.: 72401  
 Name: Inoculating Fluid -A (IF-A)  
 Lot No.: 21APR081  
 Expiration Date: 08 APR 2022

QUALITY PROFILE

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
NO GROWTH PROMOTION REQUIRED		

Final pH: 7.1 ± 0.2 at 25 °C.  
 Color: Light Green  
 Unit: tube  
 Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

  
 APPROVAL: QUALITY ASSURANCE

05 May 2021  
 DATE