

21124 Cabot Blvd. Hayward, California 94545 510-785-2564 Telephone 800-284-4949 Orders 510-782-4639 Fax www.biolog.com

CERTIFICATE OF PERFORMANCE

Pr	odu	ct	No	•

72401

Name:

Inoculating Fluid -A (IF-A)

Lot No .:

21MAY181

Expiration Date:

18 MAY 2022

QUALITY PROFILE

Challenge organisms

Source

Growth Result

NO GROWTH PROMOTION REQUIRED

Final pH: 7.1 ± 0.2 at 25 °C.

Color:

Light Green

Unit:

tube

Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C - 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

APPROVAL: QUALITY ASSURANCE

9 June 2021