



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF PERFORMANCE

Product No.: 72401  
Name: Inoculating Fluid -A (IF-A)  
Lot No.: 21NOV021  
Expiration Date: 02 NOV 2022

#### QUALITY PROFILE

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

NO GROWTH PROMOTION REQUIRED

Final pH:  $7.1 \pm 0.2$  at 25 °C.  
Color: Light Green  
Unit: tube  
Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

#### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

  
APPROVAL: QUALITY ASSURANCE

29 Nov 2021  
DATE