



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF ANALYSIS

Product No.: 3426  
Name: FF Turbidity Standard  
Lot No.: F020722  
Expiration Date: 07 NOV 2023

---

#### QUALITY PROFILE

---

##### Results

Standard check:  
75.0 ± 0.5 % transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

---

#### STORAGE RECOMMENDATIONS

Store in the dark at 20° - 26°C.

---

#### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

---

  
APPROVAL: QUALITY ASSURANCE

16 Feb 2022  
DATE