



21124 Cabot Blvd.
 Hayward, California 94545
 510-785-2564 Telephone
 800-284-4949 Orders
 510-782-4639 Fax
 www.biolog.com

CERTIFICATE OF PERFORMANCE

Product No.: 72401
 Name: Inoculating Fluid -A (IF-A)
 Lot No.: 22FEB031
 Expiration Date: 03 FEB 2023

QUALITY PROFILE

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
NO GROWTH PROMOTION REQUIRED		

Final pH: 7.1 ± 0.2 at 25 °C.
 Color: Light Green
 Unit: tube
 Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.


 APPROVAL: QUALITY ASSURANCE

22 Feb 2022
 DATE