

Product: Biolog Pipette Tips, Sterile Filtered - 1250 µL
Packaging: VistaRak, 192/Rack, 960/CS
Catalog Number: 3203
Lot Number: 22052S
Manufacture Date: May 09, 2022
Expiry Date: May 08, 2025

Tip material: Polypropylene

Filter Material: Polyethylene

Sterilization: < 10⁻⁶ SAL

Pre-sterilization is by electron beam radiation with post-sterilization verification. This lot has been processed and tested in accordance with ISO 11137-2. Product samples are bio-burdened.

RNase/DNase: RNase < 10⁻⁹ Kunitz units/µL, DNase < 10⁻⁷ Kunitz units/µL

This product has been tested and is free of any detectable RNase and DNase.

Endotoxin < 0.06 EU/mL

This product has been tested and meets the USP <85> and ANSI/AAMI ST72:2011 guidelines and FDA regulations of endotoxin limit for medical devices at 0.06 EU/mL.

Animal Content:

Resins, inks and packaging materials are not of animal origin and do not contain, make use of, or involve, at any point in the manufacturing process, raw materials of animal origin, including animal proteins.

VistaLab Technologies, Inc. is registered ISO 9001:2015

Signed on behalf of the Quality Department:



Jeff Calhoun
V.P. Product Development