



## Certificate of Performance

Catalog No.: 72268  
Product Name: PM IF-0a GN/GP Base (1.2X)  
Lot No.: 2208081  
Expiration Date: 08 AUG 2024

### Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

NO GROWTH PROMOTION REQUIRED


Final pH: 7.1 ± 0.2 at 25 °C.  
Color: Colorless, clear liquid  
Unit: Bottle  
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

### Intended Use

For research use only.

  
Quality Assurance Approval

  
Date



21124 Cabot Blvd  
Hayward, CA 94545 USA

Tel: +510 785 2564  
Fax: +510 782 4639

www.biolog.com  
Effective Date: 12/20/2019

CA72268  
Rev C

Document Number: QC72230.F1  
 Revision Number: A  
 Effective Date: 30 August 2017  
 Page: 1 of 1  
 Date Printed: Monday, 28 August 2017



*Note: Printed documents are for reference only and may be outdated. Refer to controlled version to verify latest release.*

**Document Name: PM IF Quality Control Form**

**Section 1: Product Information**

Part No.:	72268
Product Name:	IF - 0a
Lot No.:	2208081
Expiration Date:	08 AUG 2024
Setup Technician:	Antonio Munguia
Setup Date:	08-17-2022

**Section 2: Quality Profile**

Contamination Check: G = Growth, NG = No Growth

Tube	Tube Type	Day 1	Day 2	Day 3	Day 4
1	26 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG
2	26 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG
3	26 °C, C	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG	<input type="checkbox"/> G <input type="checkbox"/> NG
4	30 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG
5	30 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG
6	30 °C, C	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG
7	36 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG
8	36 °C, 1:2	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG
9	36 °C, C	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG	<input type="checkbox"/> G <input checked="" type="checkbox"/> NG

Appearance and pH:

Appearance:	Clear, colorless liquid
pH:	7.05

**Section 3: Results**

Result (Pass/Fail):	Pass	
Technician Name:	Antonio Munguia	
Signature:	<i>[Signature]</i>	
Date:	08/22/22	
QA Approval:	<i>[Signature]</i>	
Date:	08/22/2022	

*This document contains confidential information owned by Biolog, Inc. Neither this document nor the information herein may be reproduced, used or disclosed to any unauthorized third party without Biolog approval.*

PM IF-10a GN/GP BASE (1.2X)

Not For In Vitro Diagnostic Use

LOT SAMPLES TESTED AND

MET BIOLOG STANDARDS

PROD # 72264 125 ml BOTTLE

STORE @ 2 to 8°C PROTECT FROM LIGHT

LOT 2208101 EXP 2023-08-10

BIOLOG Hayward, CA 94545