



21124 Cabot Blvd.  
 Hayward, California 94545  
 510-785-2564 Telephone  
 800-284-4949 Orders  
 510-782-4639 Fax  
 www.biolog.com

**CERTIFICATE OF PERFORMANCE**

Product No.: 72402  
 Name: Inoculating Fluid-B (IF-B)  
 Lot No.: 23JAN051  
 Expiration Date: 05 JAN 2025

**QUALITY PROFILE**

| <u>Challenge organisms</u>   | <u>Source</u> | <u>Growth Result</u> |
|------------------------------|---------------|----------------------|
| NO GROWTH PROMOTION REQUIRED |               |                      |

Final pH: 6.9 ± 0.4 at 25 °C.  
 Color: Light Yellow  
 Unit: tube  
 Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C - 35 °C for 48 hours and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

**INTENDED USE**

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

David F...  
 APPROVAL: QUALITY ASSURANCE

01/16/2023  
 DATE