



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF ANALYSIS

Product No.: 3440  
Name: 65% Turbidity Standard  
Lot No.: A031423  
Expiration Date: 14 DEC 2024

---

### QUALITY PROFILE

---

	Results
Standard check: 65.0 ± 0.5 % transmittance	Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

---

### STORAGE RECOMMENDATIONS

Store in the dark at 20° - 26°C.


---

### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

---

  
APPROVAL: QUALITY ASSURANCE

  
DATE