



## Certificate of Performance

Catalog No.: 72007  
Product Name: AN Inoculating Fluid (AN-IF)  
Lot No.: A042023  
Expiration Date: 20 APR 2024

---

### Quality Profile

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required.

---

Final pH:  $7.1 \pm 0.1$  at 25 °C  
Color: Clear  
Unit: Tube  
Volume: 14 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 35 °C for 72 hours in CMG for sterility and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

---

### Intended Use

---

Not for human *in vitro* diagnostic use.

  
Quality Assurance Approval

05/01/2023  
Date



21124 Cabot Blvd  
Hayward, CA 94545 USA

Tel: +510 785 2564  
Fax: +510 782 4639

www.biolog.com

CA72007  
Rev. C  
Effective Date: 12/20/2019