



21124 Cabot Blvd.
Hayward, California 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF ANALYSIS

Product No.: 3414
Name: GP-COC & GP-ROD & GN-FAS Turbidity Standard
Lot No.: G 112414
Expiration Date: 24 AUG 2016

QUALITY PROFILE

	Results
Low Standard check: 20.0 ± 0.5 % transmittance	Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

STORAGE RECOMMENDATIONS

Store in the dark at 20° - 26°C.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

25 NOV 14

APPROVAL: QUALITY ASSURANCE
Daniel Rice

DATE