



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF ANALYSIS

Product No.: 3415  
Name: YT Turbidity Standard  
Lot No.: Y030216  
Expiration Date: 02 DEC 2017

---

#### QUALITY PROFILE

---

#### Results

Standard check:  
47.0 ± 0.5 % Transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

---

#### STORAGE RECOMMENDATIONS

Store in the dark at 20 °C – 26 °C

---

#### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

---

Lia Ricamonte  
APPROVAL: QUALITY ASSURANCE

04 Mar 2015  
DATE