



21124 Cabot Blvd.  
Hayward, California 94545  
510-785-2564 Telephone  
800-284-4949 Orders  
510-782-4639 Fax  
www.biolog.com

### CERTIFICATE OF ANALYSIS

Product No.: 3441  
Name: 85%T Turbidity Standard  
Lot No.: M010818  
Expiration Date: 08 OCT 2019

---

### QUALITY PROFILE

---

	Results
Standard check: 85.0 ± 0.5 % Transmittance	Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

---

### STORAGE RECOMMENDATIONS

Store in the dark at 20 °C – 26 °C.

---

### INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.

---

  
APPROVAL: QUALITY ASSURANCE

09Jan2018  
DATE