



21124 Cabot Blvd.
Hayward, California 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF ANALYSIS

Product No.: 3441
Name: 85%T Turbidity Standard
Lot No.: M090817
Expiration Date: 08 JUN 2019

QUALITY PROFILE

Results

Standard check:
85.0 ± 0.5 % Transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

STORAGE RECOMMENDATIONS

Store in the dark at 20 °C – 26 °C.

INTENDED USE

NOT FOR HUMAN IN VITRO DIAGNOSTIC USE.


APPROVAL: QUALITY ASSURANCE

14 Sep 2017
DATE