



21124 Cabot Blvd.
Hayward, CA 94545
510-785-2564 Telephone
800-284-4949 Orders
510-782-4639 Fax
www.biolog.com

CERTIFICATE OF PERFORMANCE

Catalog No.: 70101
Product Name: BUG™ Dehydrated Media
Lot No.: 703301
Expiration Date: 04/2020

QUALITY PROFILE

Quality Control Organism	Source (ATCC)	Medium	Incubation	Growth Response
<i>Bacillus badius</i>	14574(der.)	0.25% Mal	Aerobic 30 °C	Growth
<i>Bacillus subtilis</i>	6633(der.)	0.25% Mal	Aerobic 30 °C	Growth
<i>Enterococcus faecalis</i>	4083	5% Blood	Aerobic 35 °C	Growth
<i>Corynebacterium matruchotii</i>	14266	5% Blood	Aerobic 35 °C	Growth
<i>Micrococcus diversus</i>	25590	5% Blood	Aerobic 30 °C	Growth
<i>Staphylococcus aureus</i>	6538(der.)	5% Blood	Aerobic 35 °C	Growth
<i>Streptococcus pneumoniae</i>	33400	5% Blood	CO ₂ 35 °C	Growth
<i>Bordetella avium</i>	35086	5% Blood	Aerobic 35 °C	Growth
<i>Escherichia coli</i>	8739(der.)	5% Blood	Aerobic 35 °C	Growth
<i>Obesumbacterium proteus</i>	12841	5% Blood	Aerobic 35 °C	Growth
<i>Pseudomonas aeruginosa</i>	9027(der.)	5% Blood	Aerobic 30 °C	Growth
<i>Stenotrophomonas maltophilia</i>	13637	5% Blood	Aerobic 30 °C	Growth

Final pH: 7.3 ± 0.1 at 25 °C, BUG with Maltose 7.3 ± 0.1 at 25 °C, BUG with Blood

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc.

Technician: PT _____ **Completion Date:** 24MAR2017 _____

INTENDED USE

Not for human *in vitro* diagnostic use.



Quality Assurance Approval

10 May 2017

Date