



Certificate of Analysis

RPMI IF M1

Catalog Number: 72301
Lot Number: 707001
Storage: 2°C – 8°C
Date of Manufacture: July, 2017
Retest Date: July 31, 2020

Test	Specification	Result
Appearance	Clear, colorless to yellow solution	Conforms
Osmolarity	270 – 300mOsm/kg	272mOsm/kg
pH	7.00 – 7.40	7.37
Sterility		
Aerobic	Negative	Negative
Anaerobic	Negative	Negative
Endotoxin	≤ 1.0EU/mL	< 0.25EU/mL
Mycoplasma	Negative	Negative
Animal Cell Culture		
Growth Promotion	3 passages	Conforms
Viability	≥ 80%	> 85%

Approved by:

Date: July 27, 2017

John Grubb, PhD
Gojira Fine Chemicals, LLC
Quality Manager
Tel: 440-539-7033 Fax: 888-211-5523
john@gojirafc.com