

CERTIFICATE OF ANALYSIS

Catalog No.: 72231
Product Name: PM IFY-0 Base (1.2X)
Lot No.: 2311091
Expiration Date: 09 NOV 2024

Quality Profile

| <u>Challenge organisms</u> | <u>Source</u> | <u>Growth Result</u> |
|----------------------------|---------------|----------------------|
|----------------------------|---------------|----------------------|

No Growth Promotion Required

Final pH: 6.0 ± 0.2 at 25 °C.
Color: Colorless, slightly turbid liquid
Unit: Bottle
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

11/20/2023
Date

biolog

21124 Cabot Blvd.
Hayward, CA 94545

biolog.com
+1 800 284 4949

CA72231 RevD
Effective Date: 10/20/2023