

# CERTIFICATE OF ANALYSIS

Catalog No.: 72501  
Product Name: YT Inoculating Fluid  
Lot No.: 2312061  
Expiration Date: 06 DEC 2025

## Quality Profile

Parameter	Result
Appearance:	Passes test
Contamination check:	Passes test

Samples of this lot have met appearance and sterility standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.


Color: Clear, colorless liquid  
Unit: Tube  
Volume: 12.5 mL

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

  
Quality Assurance

  
Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA72501 RevC  
Effective Date: 10/20/2023