

CERTIFICATE OF ANALYSIS

Catalog No.: 72501
Product Name: YT Inoculating Fluid
Lot No.: 2402141
Expiration Date: 14 FEB 2026

Quality Profile

Parameter	Result
Appearance:	Passes test
Contamination check:	Passes test

Samples of this lot have met appearance and sterility standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Color: Clear, colorless liquid
Unit: Tube
Volume: 12.5 mL


Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only



Quality Assurance



Date

biolog

21124 Cabot Blvd.
Hayward, CA 94545

biolog.com
+1 800 284 4949

CA72501 RevC
Effective Date: 10/20/2023