

# CERTIFICATE OF ANALYSIS

**Catalog No.:** 71007  
**Product Name:** BUA™ Agar Plated Media w/ 5% Sheep's Blood  
**Lot No.:** 2402291  
**Expiration Date:** 28 FEB 2025

## Quality Profile

Quality Control Organism	Source (ATCC)	Growth Response	Correct Identification?
<i>Bifidobacterium breve</i>	15700	Growth	Yes
<i>Clostridium sordellii</i>	9714	Growth	Yes
<i>Lactobacillus casei</i>	393	Growth	Yes
<i>Micromonas micros</i>	33270	Growth	Yes

**Final pH:** 7.2 ± 0.1 at 25 °C

**Unit:** Plate

**Volume:** 23 mL

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

  
Quality Assurance

04/01/2024  
Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA71007 RevC  
Effective Date: 10/20/2023