

CERTIFICATE OF ANALYSIS

CERTIFICATE OF PERFORMANCE

Catalog No.: 72007
Product Name: AN Inoculating Fluid (AN-IF)
Lot No.: A032024
Expiration Date: 20 MAR 2025

Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required.

Final pH: 7.1 ± 0.1 at 25 °C
Color: Clear
Unit: Tube
Volume: 14 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 35 °C for 72 hours in CMG for sterility and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

04/02/2024
Date

biolog

21124 Cabot Blvd. biolog.com
Hayward, CA 94545 +1 800 284 4949

CA72007 RevD
Effective Date: 10/20/2023