

# CERTIFICATE OF ANALYSIS

**Catalog No.:** 72268  
**Product Name:** PM IF-0a GN/GP Base (1.2X)  
**Lot No.:** 2404171  
**Expiration Date:** 17 APR 2026

---

## Quality Profile

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
No Growth Promotion Required		

---

Final pH: 7.1 ± 0.2 at 25 °C.  
Color: Colorless, clear liquid  
Unit: Bottle  
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.


---

## Intended Use


Not for human *in vitro* diagnostic use

For Research Use Only

---



Quality Assurance



Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA72268 RevD  
Effective Date: 10/20/2023