

# CERTIFICATE OF ANALYSIS

Catalog No.: 72231  
Product Name: PM IFY-0 Base (1.2X)  
Lot No.: 2406131  
Expiration Date: 13 JUN 2025

---

## Quality Profile

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required

---

Final pH:	6.0 ± 0.2 at 25 °C.
Color:	Colorless, slightly turbid liquid
Unit:	Bottle
Volume:	125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

---

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

---



Quality Assurance



Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA72231 RevD  
Effective Date: 10/20/2023