

CERTIFICATE OF ANALYSIS

Catalog No.: 3426
Product Name: FF Turbidity Standard
Lot No.: F060524
Expiration Date: 05 MAR 2026

Quality Profile

Results

Standard check: 75.0 ± 0.5% transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

06/10/2024
Date