

# CERTIFICATE OF ANALYSIS

Catalog No.: 3415  
Product Name: YT Turbidity Standard  
Lot No.: Y082124  
Expiration Date: 21 MAY 2026

## Quality Profile

### Results

Standard check:  $47.0 \pm 0.5\%$  transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.


## Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

  
Quality Assurance

26 AUG 2024  
Date

**biolog**

21124 Cabot Blvd. biolog.com  
Hayward, CA 94545 +1 800 284 4949

CA3415 RevB  
Effective Date: 10/20/2023