

CERTIFICATE OF ANALYSIS

Catalog No.: 3440
Product Name: 65% Turbidity Standard
Lot No.: A102423
Expiration Date: 24 JUL 2025

Quality Profile

Results

Standard check:
65.0 ± 0.5% transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

11/01/2023
Date

biolog

21124 Cabot Blvd. biolog.com
Hayward, CA 94545 +1 800 284 4949

CA3440 RevB
Effective Date: 10/20/2023