

CERTIFICATE OF ANALYSIS

Catalog No.: 3441
Product Name: 85% Turbidity Standard
Lot No.: M123124
Expiration Date: 31 SEP 2026

Quality Profile

Results

Standard check:
85.0 ± 0.5% transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

02 JAN 2025
Date

biolog

21124 Cabot Blvd. biolog.com
Hayward, CA 94545 +1 800 284 4949

CA3441 RevB
Effective Date: 10/20/2023