

# CERTIFICATE OF ANALYSIS

Catalog No.: 72268  
Product Name: PM IF-0a GN/GP Base (1.2X)  
Lot No.: 2411211  
Expiration Date: 21 NOV 2026

## Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required

Final pH: 7.1 ± 0.2 at 25 °C.  
Color: Colorless, clear liquid  
Unit: Bottle  
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

  
Quality Assurance

23 DEC 2024  
Date

**biolog**

21124 Cabot Blvd. biolog.com  
Hayward, CA 94545 +1 800 284 4949

CA72268 RevD  
Effective Date: 10/20/2023