

# CERTIFICATE OF ANALYSIS

Catalog No.: 3427  
Product Name: AN Turbidity Standard  
Lot No.: A121924  
Expiration Date: 19 SEP 2026

---

## Quality Profile

---

### Results

Standard check:  $65.0 \pm 0.5\%$  transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.

---

### Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

---

### Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

---



Quality Assurance

27 DEC 2021

Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA3427 RevB  
Effective Date: 10/20/2023