

CERTIFICATE OF ANALYSIS

Product No.: 72401
Name: Inoculating Fluid-A (IF-A)
Lot No.: 25JAN071
Expiration Date: 07 JAN 2027

Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required

Final pH: 6.9 ± 0.4 at 25 °C.
Color: Light Green
Unit: Tube
Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at the time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

72 JAN 2025
Date

biolog

21124 Cabot Blvd.
Hayward, CA 94545

biolog.com
+1 800 284 4949

CA72401 RevD
Effective Date: 10/20/2023