

# CERTIFICATE OF ANALYSIS

Catalog No.: 3441  
Product Name: 85% Turbidity Standard  
Lot No.: M123124  
Expiration Date: 30 SEP 2026

---

## Quality Profile

---

### Results

Standard check: 85.0 ± 0.5% transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter. Reference turbidimeters are calibrated with NIST traceable standards.

---

## Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

---

## Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

---

  
Quality Assurance

27 FEB 2025  
Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA3441 RevB  
Effective Date: 10/20/2023