

CERTIFICATE OF ANALYSIS

Catalog No.: 3427
Product Name: AN Turbidity Standard
Lot No.: A031425
Expiration Date: 14 DEC 2026

Quality Profile

Results

Standard check: $65.0 \pm 0.5\%$ transmittance Passes test

Note: The % transmittance values are obtained from specific reference turbidimeters. These values may differ from turbidimeter to turbidimeter.


Storage Recommendations:

Store in the dark between 20 °C and 26 °C.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only



Quality Assurance



Date

biolog

21124 Cabot Blvd.
Hayward, CA 94545

biolog.com
+1 800 284 4949

CA3427 RevB
Effective Date: 10/20/2023