

CERTIFICATE OF ANALYSIS

Catalog No.: 72264
Product Name: PM IF-10a GP Base (1.2X)
Lot No.: 2504231
Expiration Date: 23 APR 2026

Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required


Final pH:	7.1 ± 0.2 at 25 °C.
Color:	Light yellow, clear liquid
Unit:	Bottle
Volume:	125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only



Quality Assurance

15 MAY 2025

Date

biolog

21124 Cabot Blvd.
Hayward, CA 94545

biolog.com
+1 800 284 4949

CA72264 RevD
Effective Date: 10/20/2023