

CERTIFICATE OF ANALYSIS

Catalog No.: 72268
Product Name: PM IF-0a GN/GP Base (1.2X)
Lot No.: 2506181
Expiration Date: 18 JUN 2027

Quality Profile

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
No Growth Promotion Required		

Final pH: 7.1 ± 0.2 at 25 °C.
Color: Colorless, clear liquid
Unit: Bottle
Volume: 125 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only


Quality Assurance

07 JUL 2025
Date

biolog

21124 Cabot Blvd. biolog.com
Hayward, CA 94545 +1800 284 4949

CA72268 RevD
Effective Date: 10/20/2023